Technical Datasheet

Date: 02.06.22

Page 1 of 1



		EN ISO 13485 certified	
Article - No:	04-032-0400	Revision: 5	
Remarks	3-linkages	single use only	
	Specifica	ation	
Product	PCR microcentrifuge tube PP, 0,2ml, 8-strips, individually attached flat & transparent caps, 3-linkages, highly transparent, np pcr ready, max. RCF 20.000g, autocl. up to 121°C		
Colour	highly transparent		
Nominal Capacity	0,2ml		
Material Characteristics	polypropylene (PP)		
Temperature range	-80°C - 121°C		
Autoclavability	autocl. up to 121°C		
Centrifugation	max. RCF 20.000g		
Product Characteristics	biologically inert		
	physiologically harmless		
	no heavy metals according to EC/94/62		
	Food contact: raw material in compliance with regulation EC 1935/2004		
Purity Grade	free of PCR inhibitors		
	free of human DNA und RNA		
	free of DNase and RNase		
	free of pyrogens		
	free of endotoxins		
Shelf - Life	not limited at controlled storage conditions		
Packaging	pcs per inner cs: 120		
	pcs per carton: 1.200		
	pcs per pallet: 57.600		
Storage	dry condition: 40%-60% rel. humidity		
	bags / cases orig. closed		
	storage temp. +2°C to max. +30°C		
	pal. not stackable		
Approved by nerbe plus	quality department		

nerbe plus GmbH & Co. KG Porschestr. 25 ir D-21423 Winsen/Luhe w Tel: +49 (0)4171 7849 0 H Fax:+49 (0)4171 6 45 45 U

info@nerbe-plus.de www.nerbe-plus.de HRA Lüneburg 203240 US1.-ID: DE 32583 0652 Pers. haftende Gesellschafterin: NERBE R&D GmbH, Winsen/Luhe GF: Philipp J.F. Nerbe HRB Lüneburg 208826 Deutsche Bank AG Kto. 0204 41200 IBAN: DE41 2407 0075 0020 4412 00 SWIFT / BIC DEUTDE2H240 Hamburger Sparkasse Kto. 1387 140831 IBAN: DE56 2005 0550 1387 1408 31 SWIFT / BIC HASPDEHHXXX

