Technical Datasheet

Date: 22.06.20

Page 1 of 1



		EN ISO 13485 certified
Article - No:	12-001-0100	Revision: 1
Remarks	Universal fit	single use only
Specification		
Product	GTS-pipette, Single-channel, variable, 10-100µl, color code yellow, CE	
Classification	CE	
Colour	color code yellow	
Nominal Capacity	10-100µl	
Purity Grade	manufactured in clean rooms of ISO Class 8 or better (DIN EN ISO 14644-1)	
Shelf - Life	not limited at controlled storage conditions	
Approved by nerbe plus quality department		

00-000-0001/06.2020

nerbe plus GmbH & Co. KG Porschestr. 25 D-21423 Winsen/Luhe Tel: +49 (0)4171 7849 0 Fax:+49 (0)4171 6 45 45

info@nerbe-plus.de www.nerbe-plus.de HRA Lüneburg 203240 USt.-ID: DE 32583 0652

Pers. haftende Gesellschafterin: NERBE R&D GmbH, Winsen/Luhe GF: Philipp J.F. Nerbe HRB Lüneburg 208826 Deutsche Bank AG Kto. 0204 41200 IBAN: DE41 2407 0075 0020 4412 00 SWIFT / BIC DEUTDE2H240

Hamburger Sparkasse Kto. 1387 140831 IBAN: DE56 2005 0550 1387 1408 31 SWIFT / BIC HASPDEHHXXX

